



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

ELAP Certificates 1551, 2728, and 2922

22 August 2016

West Bay Sanitary District

Attn: John Simonetti

500 Laurel Street

Menlo Park, CA 94025

RE: WQ Monitoring

Work Order: 16H0614

Enclosed are the results of analyses for samples received by the laboratory on 08/04/16 21:00. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Jeanette L. Poplin For Robbie C. Phillips

Project Manager



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
G13223 Grab 080316-1500	16H0614-01	Water	08/03/16 15:00	08/04/16 21:00



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

	Result	Reporting Limit	Dilution	Batch	Prepared	Analyzed	Method	Note
G13223 Grab 080316-1500 (16H0614-01)								
Sample Type: Water								
Sampled: 08/03/16 15:00								
Conventional Chemistry Parameters by APHA/EPA Methods								
Hydrogen sulfide	0.0308 mg/L	0.0200	1	AH63759	08/17/16 13:11	08/17/16 13:12	SM4500-S2 H	
pH	7.17 pH Units	1.68	1	AH63381	08/05/16 16:00	08/05/16 17:00	SM4500-H+ B	T-14
Sulfide, dissolved	ND mg/L	0.20	2	AH63505	08/05/16 13:30	08/10/16 15:30	SM4500SD	T-13
Volatile Organic Compounds by EPA Method 624								
Acetone	64 ug/L	5.0	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Benzene	ND ug/L	0.30	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Bromodichloromethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Bromoform	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Bromomethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Carbon tetrachloride	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Chlorobenzene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Chloroethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Chloroform	3.4 ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Chloromethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Dibromochloromethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,2-Dibromoethane (EDB)	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,2-Dichlorobenzene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,3-Dichlorobenzene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,4-Dichlorobenzene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,1-Dichloroethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,2-Dichloroethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,1-Dichloroethene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
cis-1,2-Dichloroethene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
trans-1,2-Dichloroethene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,2-Dichloropropane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
cis-1,3-Dichloropropene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
trans-1,3-Dichloropropene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,3-Dichloropropene (total)	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Ethylbenzene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Methyl ethyl ketone	4.6 ug/L	1.0	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Methyl isobutyl ketone	ND ug/L	1.0	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Methylene chloride	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Styrene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,1,2,2-Tetrachloroethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Tetrachloroethene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Toluene	1.5 ug/L	0.30	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
1,1,1-Trichloroethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

	Result	Reporting Limit	Dilution	Batch	Prepared	Analyzed	Method	Note
G13223 Grab 080316-1500 (16H0614-01)		Sample Type: Water			Sampled: 08/03/16 15:00			
Volatile Organic Compounds by EPA Method 624 (cont'd)								
1,1,2-Trichloroethane	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Trichloroethene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Vinyl chloride	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
o-Xylene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
m,p-Xylene	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Xylenes (total)	ND ug/L	0.50	1	AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Surrogate: Bromofluorobenzene	100 %	81-135		AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Surrogate: Dibromofluoromethane	83.8 %	46-130		AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	
Surrogate: Toluene-d8	91.4 %	59-132		AH63415	08/08/16 11:00	08/09/16 07:11	EPA 624	



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Conventional Chemistry Parameters by APHA/EPA Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63505 - General Preparation										
Blank (AH63505-BLK1)				Prepared: 08/03/16 Analyzed: 08/10/16						
Sulfide, dissolved	ND	0.10	mg/L							
LCS (AH63505-BS1)				Prepared & Analyzed: 08/10/16						
Sulfide, dissolved	0.495	0.10	mg/L	0.492		101	80-120			
Duplicate (AH63505-DUP1)				Source: 16H0926-01		Prepared & Analyzed: 08/10/16				
Sulfide, dissolved	ND	0.10	mg/L		ND				15	
Matrix Spike (AH63505-MS1)				Source: 16H0926-01		Prepared & Analyzed: 08/10/16				
Sulfide, dissolved	0.258	0.10	mg/L	0.492	ND	52.4	80-120			QM-05
Matrix Spike Dup (AH63505-MSD1)				Source: 16H0926-01		Prepared & Analyzed: 08/10/16				
Sulfide, dissolved	0.258	0.10	mg/L	0.492	ND	52.4	80-120	0.00	15	QM-05



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Volatile Organic Compounds by EPA Method 624 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63415 - VOAs in Water GCMS										
Blank (AH63415-BLK1)				Prepared & Analyzed: 08/08/16						
Acetone	ND	5.0	ug/L							
Benzene	ND	0.30	ug/L							
Bromodichloromethane	ND	0.50	ug/L							
Bromoform	ND	0.50	ug/L							
Bromomethane	ND	0.50	ug/L							
Carbon tetrachloride	ND	0.50	ug/L							
Chlorobenzene	ND	0.50	ug/L							
Chloroethane	ND	0.50	ug/L							
Chloroform	ND	0.50	ug/L							
Chloromethane	ND	0.50	ug/L							
Dibromochloromethane	ND	0.50	ug/L							
1,2-Dibromoethane (EDB)	ND	0.50	ug/L							
1,2-Dichlorobenzene	ND	0.50	ug/L							
1,3-Dichlorobenzene	ND	0.50	ug/L							
1,4-Dichlorobenzene	ND	0.50	ug/L							
1,1-Dichloroethane	ND	0.50	ug/L							
1,2-Dichloroethane	ND	0.50	ug/L							
1,1-Dichloroethene	ND	0.50	ug/L							
cis-1,2-Dichloroethene	ND	0.50	ug/L							
trans-1,2-Dichloroethene	ND	0.50	ug/L							
1,2-Dichloropropane	ND	0.50	ug/L							
cis-1,3-Dichloropropene	ND	0.50	ug/L							
trans-1,3-Dichloropropene	ND	0.50	ug/L							
1,3-Dichloropropene (total)	ND	0.50	ug/L							
Ethylbenzene	ND	0.50	ug/L							
Methyl ethyl ketone	ND	1.0	ug/L							
Methyl isobutyl ketone	ND	1.0	ug/L							
Methylene chloride	ND	0.50	ug/L							
Styrene	ND	0.50	ug/L							
1,1,2,2-Tetrachloroethane	ND	0.50	ug/L							
Tetrachloroethene	ND	0.50	ug/L							
Toluene	ND	0.30	ug/L							
1,1,1-Trichloroethane	ND	0.50	ug/L							

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Volatile Organic Compounds by EPA Method 624 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63415 - VOAs in Water GCMS										
Blank (AH63415-BLK1)				Prepared & Analyzed: 08/08/16						
1,1,2-Trichloroethane	ND	0.50	ug/L							
Trichloroethene	ND	0.50	ug/L							
Vinyl chloride	ND	0.50	ug/L							
o-Xylene	ND	0.50	ug/L							
m,p-Xylene	ND	0.50	ug/L							
Xylenes (total)	ND	0.50	ug/L							
Surrogate: Bromofluorobenzene	24.1		ug/L	25.0		96.4	81-135			
Surrogate: Dibromofluoromethane	20.6		ug/L	25.0		82.6	46-130			
Surrogate: Toluene-d8	23.2		ug/L	25.0		92.6	59-132			
LCS (AH63415-BS1)				Prepared & Analyzed: 08/08/16						
Acetone	62.9	5.0	ug/L	80.0		78.6	48-124			
Benzene	20.5	0.30	ug/L	20.0		102	82-122			
Bromodichloromethane	19.4	0.50	ug/L	20.0		97.2	86-135			
Bromoform	19.4	0.50	ug/L	20.0		97.0	76-144			
Bromomethane	19.9	0.50	ug/L	20.0		99.4	69-145			
Carbon tetrachloride	20.4	0.50	ug/L	20.0		102	77-134			
Chlorobenzene	19.8	0.50	ug/L	20.0		98.8	84-119			
Chloroethane	19.2	0.50	ug/L	20.0		96.0	68-133			
Chloroform	19.2	0.50	ug/L	20.0		96.0	81-122			
Chloromethane	18.6	0.50	ug/L	20.0		93.0	63-129			
Dibromochloromethane	19.9	0.50	ug/L	20.0		99.6	83-135			
1,2-Dibromoethane (EDB)	20.0	0.50	ug/L	20.0		100	80-120			
1,2-Dichlorobenzene	20.0	0.50	ug/L	20.0		100	84-121			
1,3-Dichlorobenzene	20.5	0.50	ug/L	20.0		103	80-120			
1,4-Dichlorobenzene	19.7	0.50	ug/L	20.0		98.3	84-120			
1,1-Dichloroethane	19.6	0.50	ug/L	20.0		98.0	81-126			
1,2-Dichloroethane	18.9	0.50	ug/L	20.0		94.4	77-117			
1,1-Dichloroethene	19.7	0.50	ug/L	20.0		98.3	71-151			
cis-1,2-Dichloroethene	19.5	0.50	ug/L	20.0		97.4	84-131			
trans-1,2-Dichloroethene	19.3	0.50	ug/L	20.0		96.4	79-128			
1,2-Dichloropropane	20.1	0.50	ug/L	20.0		100	82-125			
cis-1,3-Dichloropropene	21.1	0.50	ug/L	20.0		105	83-128			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Volatile Organic Compounds by EPA Method 624 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63415 - VOAs in Water GCMS										
LCS (AH63415-BS1)				Prepared & Analyzed: 08/08/16						
trans-1,3-Dichloropropene	20.1	0.50	ug/L	20.0		101	67-129			
Ethylbenzene	21.2	0.50	ug/L	20.0		106	84-124			
Methyl ethyl ketone	33.7	1.0	ug/L	40.0		84.4	58-157			
Methyl isobutyl ketone	37.9	1.0	ug/L	40.0		94.8	70-130			
Methylene chloride	20.3	0.50	ug/L	20.0		101	72-132			
Styrene	22.9	0.50	ug/L	20.0		114	80-125			
1,1,2,2-Tetrachloroethane	18.6	0.50	ug/L	20.0		93.0	84-115			
Tetrachloroethene	21.0	0.50	ug/L	20.0		105	56-156			
Toluene	21.3	0.30	ug/L	20.0		107	76-137			
1,1,1-Trichloroethane	20.7	0.50	ug/L	20.0		103	70-130			
1,1,2-Trichloroethane	19.5	0.50	ug/L	20.0		97.7	83-122			
Trichloroethene	20.7	0.50	ug/L	20.0		104	84-123			
Vinyl chloride	19.6	0.50	ug/L	20.0		97.8	70-130			
o-Xylene	21.9	0.50	ug/L	20.0		110	80-126			
m,p-Xylene	44.4	0.50	ug/L	40.0		111	81-124			
Xylenes (total)	66.3	0.50	ug/L	60.0		111	81-126			
Surrogate: Bromofluorobenzene	24.9		ug/L	25.0		99.8	81-135			
Surrogate: Dibromofluoromethane	20.7		ug/L	25.0		82.8	46-130			
Surrogate: Toluene-d8	22.5		ug/L	25.0		90.2	59-132			
LCS Dup (AH63415-BS1)				Prepared & Analyzed: 08/08/16						
Acetone	66.8	5.0	ug/L	80.0		83.5	48-124	6.08	25	
Benzene	20.1	0.30	ug/L	20.0		101	82-122	1.87	25	
Bromodichloromethane	20.0	0.50	ug/L	20.0		99.8	86-135	2.54	25	
Bromoform	20.0	0.50	ug/L	20.0		99.8	76-144	2.85	25	
Bromomethane	20.5	0.50	ug/L	20.0		103	69-145	3.17	25	
Carbon tetrachloride	20.3	0.50	ug/L	20.0		101	77-134	0.737	25	
Chlorobenzene	19.4	0.50	ug/L	20.0		97.2	84-119	1.73	25	
Chloroethane	18.8	0.50	ug/L	20.0		93.8	68-133	2.42	25	
Chloroform	19.3	0.50	ug/L	20.0		96.6	81-122	0.519	25	
Chloromethane	18.1	0.50	ug/L	20.0		90.3	63-129	3.00	25	
Dibromochloromethane	19.7	0.50	ug/L	20.0		98.7	83-135	0.908	25	
1,2-Dibromoethane (EDB)	20.2	0.50	ug/L	20.0		101	80-120	1.14	25	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Volatile Organic Compounds by EPA Method 624 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63415 - VOAs in Water GCMS										
LCS Dup (AH63415-BSD1)				Prepared & Analyzed: 08/08/16						
1,2-Dichlorobenzene	20.3	0.50	ug/L	20.0		101	84-121	1.34	25	
1,3-Dichlorobenzene	20.5	0.50	ug/L	20.0		102	80-120	0.0975	25	
1,4-Dichlorobenzene	19.3	0.50	ug/L	20.0		96.6	84-120	1.80	25	
1,1-Dichloroethane	19.8	0.50	ug/L	20.0		98.8	81-126	0.864	25	
1,2-Dichloroethane	19.1	0.50	ug/L	20.0		95.4	77-117	1.05	25	
1,1-Dichloroethene	19.7	0.50	ug/L	20.0		98.3	71-151	0.00	25	
cis-1,2-Dichloroethene	19.6	0.50	ug/L	20.0		97.9	84-131	0.461	25	
trans-1,2-Dichloroethene	19.6	0.50	ug/L	20.0		98.2	79-128	1.80	25	
1,2-Dichloropropane	20.1	0.50	ug/L	20.0		100	82-125	0.0498	25	
cis-1,3-Dichloropropene	21.2	0.50	ug/L	20.0		106	83-128	0.473	25	
trans-1,3-Dichloropropene	20.4	0.50	ug/L	20.0		102	67-129	1.48	25	
Ethylbenzene	21.1	0.50	ug/L	20.0		105	84-124	0.662	25	
Methyl ethyl ketone	36.4	1.0	ug/L	40.0		91.0	58-157	7.53	25	
Methyl isobutyl ketone	40.7	1.0	ug/L	40.0		102	70-130	7.02	25	
Methylene chloride	19.8	0.50	ug/L	20.0		99.2	72-132	2.15	25	
Styrene	22.7	0.50	ug/L	20.0		113	80-125	0.922	25	
1,1,2,2-Tetrachloroethane	19.8	0.50	ug/L	20.0		99.1	84-115	6.30	25	
Tetrachloroethene	20.6	0.50	ug/L	20.0		103	56-156	2.36	25	
Toluene	21.1	0.30	ug/L	20.0		106	76-137	1.04	25	
1,1,2-Trichloroethane	20.0	0.50	ug/L	20.0		100	83-122	2.43	25	
1,1,1-Trichloroethane	20.7	0.50	ug/L	20.0		103	70-130	0.0483	25	
Trichloroethene	20.2	0.50	ug/L	20.0		101	84-123	2.39	25	
Vinyl chloride	18.5	0.50	ug/L	20.0		92.4	70-130	5.78	25	
o-Xylene	22.0	0.50	ug/L	20.0		110	80-126	0.410	25	
m,p-Xylene	43.7	0.50	ug/L	40.0		109	81-124	1.63	25	
Xylenes (total)	65.7	0.50	ug/L	60.0		110	81-126	0.954	25	
Surrogate: Bromofluorobenzene	24.9		ug/L	25.0		99.7	81-135			
Surrogate: Dibromofluoromethane	20.6		ug/L	25.0		82.6	46-130			
Surrogate: Toluene-d8	22.5		ug/L	25.0		90.1	59-132			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Volatile Organic Compounds by EPA Method 624 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63415 - VOAs in Water GCMS										
Matrix Spike (AH63415-MS1)	Source: 16H0424-12			Prepared: 08/08/16 Analyzed: 08/09/16						
Acetone	64.3	5.0	ug/L	80.0	ND	80.4	32-164			
Benzene	21.3	0.30	ug/L	20.0	ND	107	58-139			
Bromodichloromethane	20.7	0.50	ug/L	20.0	ND	103	62-140			
Bromoform	21.1	0.50	ug/L	20.0	ND	106	47-165			
Bromomethane	2.42	0.50	ug/L	20.0	ND	12.1	30-163			QM-05
Carbon tetrachloride	21.9	0.50	ug/L	20.0	ND	110	65-153			
Chlorobenzene	20.7	0.50	ug/L	20.0	ND	104	58-137			
Chloroethane	19.0	0.50	ug/L	20.0	ND	95.2	59-141			
Chloroform	23.6	0.50	ug/L	20.0	3.22	102	36-151			
Chloromethane	16.5	0.50	ug/L	20.0	ND	82.7	69-149			
Dibromochloromethane	21.0	0.50	ug/L	20.0	ND	105	54-157			
1,2-Dibromoethane (EDB)	20.6	0.50	ug/L	20.0	ND	103	40-147			
1,2-Dichlorobenzene	20.0	0.50	ug/L	20.0	ND	100	57-137			
1,3-Dichlorobenzene	21.6	0.50	ug/L	20.0	ND	108	54-137			
1,4-Dichlorobenzene	19.5	0.50	ug/L	20.0	ND	97.3	61-131			
1,1-Dichloroethane	20.6	0.50	ug/L	20.0	ND	103	64-136			
1,2-Dichloroethane	19.4	0.50	ug/L	20.0	ND	97.0	58-133			
1,1-Dichloroethene	20.8	0.50	ug/L	20.0	ND	104	45-178			
cis-1,2-Dichloroethene	20.8	0.50	ug/L	20.0	ND	104	66-141			
trans-1,2-Dichloroethene	20.6	0.50	ug/L	20.0	ND	103	59-151			
1,2-Dichloropropane	20.6	0.50	ug/L	20.0	ND	103	62-128			
cis-1,3-Dichloropropene	18.1	0.50	ug/L	20.0	ND	90.4	50-140			
trans-1,3-Dichloropropene	19.1	0.50	ug/L	20.0	ND	95.4	40-144			
Ethylbenzene	22.1	0.50	ug/L	20.0	ND	111	59-147			
Methyl ethyl ketone	35.2	1.0	ug/L	40.0	ND	88.1	62-126			
Methyl isobutyl ketone	42.1	1.0	ug/L	40.0	ND	105	66-127			
Methylene chloride	19.5	0.50	ug/L	20.0	ND	97.6	43-143			
Styrene	23.9	0.50	ug/L	20.0	ND	119	51-157			
1,1,2,2-Tetrachloroethane	20.3	0.50	ug/L	20.0	ND	101	73-127			
Tetrachloroethene	22.3	0.50	ug/L	20.0	ND	111	49-148			
Toluene	23.0	0.30	ug/L	20.0	0.600	112	59-147			
1,1,1-Trichloroethane	21.5	0.50	ug/L	20.0	ND	107	62-144			
1,1,2-Trichloroethane	20.6	0.50	ug/L	20.0	ND	103	62-129			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Volatile Organic Compounds by EPA Method 624 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63415 - VOAs in Water GCMS										
Matrix Spike (AH63415-MS1)		Source: 16H0424-12		Prepared: 08/08/16		Analyzed: 08/09/16				
Trichloroethene	21.4	0.50	ug/L	20.0	ND	107	58-140			
Vinyl chloride	17.2	0.50	ug/L	20.0	ND	86.0	53-160			
o-Xylene	23.0	0.50	ug/L	20.0	ND	115	55-148			
m,p-Xylene	46.5	0.50	ug/L	40.0	ND	116	53-147			
Xylenes (total)	69.5	0.50	ug/L	60.0	ND	116	49-153			
Surrogate: Bromofluorobenzene	25.6		ug/L	25.0		103	81-135			
Surrogate: Dibromofluoromethane	20.5		ug/L	25.0		82.1	46-130			
Surrogate: Toluene-d8	22.4		ug/L	25.0		89.6	59-132			
Matrix Spike Dup (AH63415-MSD1)		Source: 16H0424-12		Prepared: 08/08/16		Analyzed: 08/09/16				
Acetone	69.2	5.0	ug/L	80.0	ND	86.5	32-164	7.32	25	
Benzene	21.9	0.30	ug/L	20.0	ND	109	58-139	2.45	25	
Bromodichloromethane	20.9	0.50	ug/L	20.0	ND	105	62-140	1.20	25	
Bromoform	21.3	0.50	ug/L	20.0	ND	106	47-165	0.755	25	
Bromomethane	4.18	0.50	ug/L	20.0	ND	20.9	30-163	53.3	25	QM-05, QM-08
Carbon tetrachloride	22.2	0.50	ug/L	20.0	ND	111	65-153	1.31	25	
Chlorobenzene	21.0	0.50	ug/L	20.0	ND	105	58-137	1.01	25	
Chloroethane	18.1	0.50	ug/L	20.0	ND	90.6	59-141	4.95	25	
Chloroform	23.8	0.50	ug/L	20.0	3.22	103	36-151	1.06	25	
Chloromethane	16.0	0.50	ug/L	20.0	ND	80.1	69-149	3.19	25	
Dibromochloromethane	21.1	0.50	ug/L	20.0	ND	106	54-157	0.475	25	
1,2-Dibromoethane (EDB)	20.2	0.50	ug/L	20.0	ND	101	40-147	1.76	25	
1,2-Dichlorobenzene	21.0	0.50	ug/L	20.0	ND	105	57-137	4.59	25	
1,3-Dichlorobenzene	22.2	0.50	ug/L	20.0	ND	111	54-137	2.64	25	
1,4-Dichlorobenzene	20.1	0.50	ug/L	20.0	ND	101	61-131	3.29	25	
1,1-Dichloroethane	21.3	0.50	ug/L	20.0	ND	107	64-136	3.63	25	
1,2-Dichloroethane	19.6	0.50	ug/L	20.0	ND	98.2	58-133	1.23	25	
1,1-Dichloroethene	21.4	0.50	ug/L	20.0	ND	107	45-178	2.75	25	
cis-1,2-Dichloroethene	21.2	0.50	ug/L	20.0	ND	106	66-141	1.95	25	
trans-1,2-Dichloroethene	21.6	0.50	ug/L	20.0	ND	108	59-151	4.69	25	
1,2-Dichloropropane	21.6	0.50	ug/L	20.0	ND	108	62-128	5.17	25	
cis-1,3-Dichloropropene	18.4	0.50	ug/L	20.0	ND	92.0	50-140	1.70	25	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Volatile Organic Compounds by EPA Method 624 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AH63415 - VOAs in Water GCMS										
Matrix Spike Dup (AH63415-MSD1)		Source: 16H0424-12		Prepared: 08/08/16 Analyzed: 08/09/16						
trans-1,3-Dichloropropene	19.5	0.50	ug/L	20.0	ND	97.5	40-144	2.12	25	
Ethylbenzene	22.7	0.50	ug/L	20.0	ND	114	59-147	2.68	25	
Methyl ethyl ketone	36.6	1.0	ug/L	40.0	ND	91.5	62-126	3.76	25	
Methyl isobutyl ketone	42.8	1.0	ug/L	40.0	ND	107	66-127	1.62	25	
Methylene chloride	21.6	0.50	ug/L	20.0	ND	108	43-143	9.88	25	
Styrene	24.2	0.50	ug/L	20.0	ND	121	51-157	1.29	25	
1,1,2,2-Tetrachloroethane	19.7	0.50	ug/L	20.0	ND	98.6	73-127	2.75	25	
Tetrachloroethene	22.4	0.50	ug/L	20.0	ND	112	49-148	0.805	25	
Toluene	22.9	0.30	ug/L	20.0	0.600	111	59-147	0.567	25	
1,1,2-Trichloroethane	20.3	0.50	ug/L	20.0	ND	101	62-129	1.71	25	
1,1,1-Trichloroethane	22.3	0.50	ug/L	20.0	ND	111	62-144	3.70	25	
Trichloroethene	21.9	0.50	ug/L	20.0	ND	109	58-140	2.03	25	
Vinyl chloride	16.3	0.50	ug/L	20.0	ND	81.4	53-160	5.56	25	
o-Xylene	23.5	0.50	ug/L	20.0	ND	117	55-148	2.11	25	
m,p-Xylene	47.6	0.50	ug/L	40.0	ND	119	53-147	2.28	25	
Xylenes (total)	71.0	0.50	ug/L	60.0	ND	118	49-153	2.22	25	
Surrogate: Bromofluorobenzene	25.1		ug/L	25.0		100	81-135			
Surrogate: Dibromofluoromethane	20.5		ug/L	25.0		82.0	46-130			
Surrogate: Toluene-d8	22.3		ug/L	25.0		89.1	59-132			



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: [none]

Reported:
08/22/16 10:40

Notes and Definitions

- QM-05 The spike recovery was outside acceptance limits for the MS and/or MSD due to matrix interference. The LCS and/or LCSD were within acceptance limits showing that the laboratory is in control and the data is acceptable.
- QM-08 The RPD was outside acceptance limits for MS/MSD, possibly due to matrix interference. The LCS and/or LCSD were within acceptance limits showing that the laboratory is in control and the data is acceptable.
- T-13 Hold time not specified in method.
- T-14 Residual chlorine, dissolved oxygen, and pH must be analyzed in the field to meet the EPA specified 15 minute hold time.
- ND Analyte NOT DETECTED at or above the reporting limit
- dry Sample results reported on a dry weight basis
- REC Recovery
- RPD Relative Percent Difference

