



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

ELAP Certificates 1551, 2728, and 2922

19 April 2016

West Bay Sanitary District

Attn: John Simonetti

500 Laurel Street

Menlo Park, CA 94025

RE: WQ Monitoring

Work Order: 16D0196

Enclosed are the results of analyses for samples received by the laboratory on 04/04/16 21:20. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Jeanette L. Poplin For Robbie C. Phillips

Project Manager



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
Oak Avenue	16D0196-01	Water	04/01/16 09:00	04/04/16 21:20



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

	Result	Reporting Limit	Dilution	Batch	Prepared	Analyzed	Method	Note
Oak Avenue (16D0196-01)								
Sample Type: Water			Sampled: 04/01/16 09:00					
Metals by EPA 200 Series Methods								
Boron	0.31 mg/L	0.050	1	AD63175	04/12/16 16:00	04/14/16 15:56	EPA 200.7	
Calcium	39 mg/L	1.0	1	AD63175	04/12/16 16:00	04/14/16 15:56	EPA 200.7	
Magnesium	23 mg/L	1.0	1	AD63175	04/12/16 16:00	04/14/16 15:56	EPA 200.7	
Sodium	73 mg/L	1.0	1	AD63175	04/12/16 16:00	04/14/16 15:56	EPA 200.7	
Conventional Chemistry Parameters by APHA/EPA Methods								
Ammonia as NH3	58 mg/L	0.50	1	AD62949	04/06/16 08:00	04/06/16 17:00	SM4500NH3C	
Bicarbonate	380 mg/L	5.0	1	AD62960	04/06/16 09:33	04/06/16 17:00	SM2320B	
Biochemical Oxygen Demand	250 mg/L	5.0	1	AD63127	04/06/16 12:00	04/11/16 13:00	SM5210B	T-1, T-16
pH	7.58 pH Units	1.68	1	AD62924	04/04/16 18:40	04/04/16 18:40	SM4500-H+ B	T-14
Specific Conductance (EC)	1100 umhos/cm	20	1	AD62910	04/05/16 17:00	04/05/16 17:00	SM2510B	
Total Dissolved Solids	510 mg/L	10	1	AD62954	04/06/16 09:00	04/12/16 08:45	SM2540C	
Total Suspended Solids	82 mg/L	1.0	1	AD62969	04/06/16 11:13	04/08/16 11:00	SM2540D	
Silica	21 mg/L	2.0	2	AD62927	04/05/16 14:15	04/05/16 16:30	SM4500-SiO2 C	
Bicarbonate Alkalinity as CaCO3	310 mg/L	5.0	1	AD62960	04/06/16 09:33	04/06/16 17:00	SM2320B	
Carbonate Alkalinity as CaCO3	ND mg/L	5.0	1	AD62960	04/06/16 09:33	04/06/16 17:00	SM2320B	
Hydroxide Alkalinity as CaCO3	ND mg/L	5.0	1	AD62960	04/06/16 09:33	04/06/16 17:00	SM2320B	
Total Alkalinity as CaCO3	310 mg/L	5.0	1	AD62960	04/06/16 09:33	04/06/16 17:00	SM2320B	
Total Kjeldahl Nitrogen	57 mg/L	1.0	1	AD62911	04/06/16 09:41	04/07/16 09:23	SM4500-Norg B	
Total Nitrogen	57 mg/L	1.0	1	AD62999	04/07/16 08:43	04/12/16 09:47	SM4500-N	
Phosphorus, total	6.0 mg/L	0.80	2	AD63089	04/11/16 08:00	04/11/16 12:48	SM4500-P E	
Miscellaneous Physical/Conventional Chemistry Parameters								
Sodium Adsorption Ratio-Adj RNA	2.70 NA		1	AD63175	04/12/16 16:00	04/19/16 07:09	SAR	



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

	Result	Reporting Limit	Dilution	Batch	Prepared	Analyzed	Method	Note
Oak Avenue (16D0196-01)		Sample Type: Water			Sampled: 04/01/16 09:00			
Anions by EPA Method 300.0								
Chloride	73 mg/L	10	20	AD62920	04/05/16 16:46	04/05/16 16:46	EPA 300.0	
Nitrate as N	ND mg/L	0.20	1	AD62920	04/05/16 17:02	04/05/16 17:02	EPA 300.0	T-2
Nitrate as NO3	ND mg/L	1.0	1	AD62920	04/05/16 17:02	04/05/16 17:02	EPA 300.0	T-2
Nitrite as N	ND mg/L	0.20	1	AD62920	04/05/16 17:02	04/05/16 17:02	EPA 300.0	T-2
Nitrite as NO2	ND mg/L	1.0	1	AD62920	04/05/16 17:02	04/05/16 17:02	EPA 300.0	T-2



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Metals by EPA 200 Series Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AD63175 - Metals Digest										
Blank (AD63175-BLK1)				Prepared: 04/12/16 Analyzed: 04/14/16						
Boron	ND	0.050	mg/L							
Calcium	ND	1.0	mg/L							
Magnesium	ND	1.0	mg/L							
Sodium	ND	1.0	mg/L							
LCS (AD63175-BS1)				Prepared: 04/12/16 Analyzed: 04/14/16						
Boron	0.386	0.050	mg/L	0.400		96.6	85-115			
Calcium	7.87	1.0	mg/L	8.00		98.4	85-115			
Magnesium	8.39	1.0	mg/L	8.00		105	85-115			
Sodium	8.35	1.0	mg/L	8.00		104	85-115			
Duplicate (AD63175-DUP1)				Source: 16D0166-01		Prepared: 04/12/16 Analyzed: 04/14/16				
Boron	0.0889	0.050	mg/L		0.0865			2.72	20	
Calcium	11.9	1.0	mg/L		11.7			1.16	20	
Magnesium	27.2	1.0	mg/L		27.6			1.55	20	
Sodium	6.84	1.0	mg/L		6.72			1.84	20	
Matrix Spike (AD63175-MS1)				Source: 16D0166-01		Prepared: 04/12/16 Analyzed: 04/14/16				
Boron	0.432	0.050	mg/L	0.400	0.0865	86.3	70-130			
Calcium	18.2	1.0	mg/L	8.00	11.7	80.9	70-130			
Magnesium	32.2	1.0	mg/L	8.00	27.6	58.2	70-130			QM-01
Sodium	13.9	1.0	mg/L	8.00	6.72	90.0	70-130			
Matrix Spike (AD63175-MS2)				Source: 16D0198-01		Prepared: 04/12/16 Analyzed: 04/14/16				
Boron	0.577	0.050	mg/L	0.400	0.197	94.8	70-130			
Calcium	88.7	1.0	mg/L	8.00	78.3	130	70-130			
Magnesium	79.4	1.0	mg/L	8.00	67.9	144	70-130			QM-4X
Sodium	51.9	1.0	mg/L	8.00	42.6	116	70-130			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Metals by EPA 200 Series Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
------------	--------	--------------------	-------	----------------	------------------	------	----------------	-----	--------------	------

Batch AD63175 - Metals Digest

Matrix Spike Dup (AD63175-MSD1)

Source: 16D0166-01

Prepared: 04/12/16 Analyzed: 04/14/16

Boron	0.462	0.050	mg/L	0.400	0.0865	93.9	70-130	6.86	20	
Calcium	19.0	1.0	mg/L	8.00	11.7	90.3	70-130	4.03	20	
Magnesium	33.7	1.0	mg/L	8.00	27.6	76.2	70-130	4.37	20	
Sodium	14.6	1.0	mg/L	8.00	6.72	98.3	70-130	4.65	20	



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Conventional Chemistry Parameters by APHA/EPA Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AD62910 - General Preparation										
Duplicate (AD62910-DUP1)		Source: 16D0236-03			Prepared & Analyzed: 04/05/16					
Specific Conductance (EC)	165	20	umhos/cm		165			0.00	5	
Batch AD62911 - General Prep										
LCS (AD62911-BS1)					Prepared: 04/05/16 Analyzed: 04/06/16					
Total Kjeldahl Nitrogen	2.95	1.0	mg/L	2.98		99.0	80-120			
LCS Dup (AD62911-BSD1)					Prepared: 04/05/16 Analyzed: 04/06/16					
Total Kjeldahl Nitrogen	2.95	1.0	mg/L	2.98		99.0	80-120	0.00	20	
Matrix Spike (AD62911-MS1)		Source: 16D0191-01			Prepared: 04/05/16 Analyzed: 04/06/16					
Total Kjeldahl Nitrogen	43.4	1.0	mg/L	22.6	24.5	83.5	75-125			
Matrix Spike Dup (AD62911-MSD1)		Source: 16D0191-01			Prepared: 04/05/16 Analyzed: 04/06/16					
Total Kjeldahl Nitrogen	43.4	1.0	mg/L	22.6	24.5	83.5	75-125	0.00	20	
Batch AD62924 - General Preparation										
Duplicate (AD62924-DUP1)		Source: 16D0229-01			Prepared & Analyzed: 04/04/16					
pH	9.00	1.68	pH Units		8.91			1.01	20	
Duplicate (AD62924-DUP2)		Source: 16D0224-01			Prepared & Analyzed: 04/04/16					
pH	8.33	1.68	pH Units		8.39			0.718	20	
Batch AD62927 - General Preparation										
Blank (AD62927-BLK1)					Prepared & Analyzed: 04/05/16					
Silica	ND	1.0	mg/L							



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Conventional Chemistry Parameters by APHA/EPA Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AD62927 - General Preparation										
LCS (AD62927-BS1)				Prepared & Analyzed: 04/05/16						
Silica	10.0	1.0	mg/L	10.0		100	85-115			
LCS Dup (AD62927-BSD1)				Prepared & Analyzed: 04/05/16						
Silica	10.6	1.0	mg/L	10.0		106	85-115	5.65	20	
Duplicate (AD62927-DUP1)				Source: 16D0230-01		Prepared & Analyzed: 04/05/16				
Silica	18.7	1.0	mg/L		18.4			1.57	20	
Matrix Spike (AD62927-MS1)				Source: 16D0230-01		Prepared & Analyzed: 04/05/16				
Silica	22.3	1.0	mg/L	5.00	18.4	78.0	80-120			QM-05
Matrix Spike Dup (AD62927-MSD1)				Source: 16D0230-01		Prepared & Analyzed: 04/05/16				
Silica	22.2	1.0	mg/L	5.00	18.4	75.1	80-120	0.656	20	QM-05
Batch AD62949 - General Preparation										
Blank (AD62949-BLK1)				Prepared & Analyzed: 04/06/16						
Ammonia as NH3	ND	0.50	mg/L							
LCS (AD62949-BS1)				Prepared & Analyzed: 04/06/16						
Ammonia as NH3	5.75	0.50	mg/L	6.10		94.2	90-110			
LCS Dup (AD62949-BSD1)				Prepared & Analyzed: 04/06/16						
Ammonia as NH3	5.85	0.50	mg/L	6.10		96.0	90-110	1.83	10	
Matrix Spike (AD62949-MS1)				Source: 16D0159-01		Prepared & Analyzed: 04/06/16				
Ammonia as NH3	12.1	0.50	mg/L	12.2	ND	99.5	85-115			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Conventional Chemistry Parameters by APHA/EPA Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
------------	--------	--------------------	-------	----------------	------------------	------	----------------	-----	--------------	------

Batch AD62949 - General Preparation

Matrix Spike Dup (AD62949-MSD1)

Source: 16D0159-01

Prepared & Analyzed: 04/06/16

Ammonia as NH ₃	11.9	0.50	mg/L	12.2	ND	97.7	85-115	1.77	20	
----------------------------	------	------	------	------	----	------	--------	------	----	--

Batch AD62954 - General Preparation

Blank (AD62954-BLK1)

Prepared: 04/06/16 Analyzed: 04/12/16

Total Dissolved Solids	ND	10	mg/L							
------------------------	----	----	------	--	--	--	--	--	--	--

Duplicate (AD62954-DUP1)

Source: 16D0213-01

Prepared: 04/06/16 Analyzed: 04/12/16

Total Dissolved Solids	600	10	mg/L		652			8.31	15	
------------------------	-----	----	------	--	-----	--	--	------	----	--

Duplicate (AD62954-DUP2)

Source: 16D0230-02

Prepared: 04/06/16 Analyzed: 04/12/16

Total Dissolved Solids	376	10	mg/L		404			7.18	15	
------------------------	-----	----	------	--	-----	--	--	------	----	--

Batch AD62960 - General Preparation

Duplicate (AD62960-DUP1)

Source: 16D0170-02

Prepared & Analyzed: 04/06/16

Bicarbonate	14.6	5.0	mg/L		14.6			0.00	20	
Hydroxide Alkalinity as CaCO ₃	ND	5.0	mg/L		ND				20	
Carbonate Alkalinity as CaCO ₃	ND	5.0	mg/L		ND				20	
Bicarbonate Alkalinity as CaCO ₃	12.0	5.0	mg/L		12.0			0.00	20	
Total Alkalinity as CaCO ₃	12.0	5.0	mg/L		12.0			0.00	20	

Batch AD62969 - General Preparation

Blank (AD62969-BLK1)

Prepared: 04/06/16 Analyzed: 04/08/16

Total Suspended Solids	ND	1.0	mg/L							
------------------------	----	-----	------	--	--	--	--	--	--	--



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Conventional Chemistry Parameters by APHA/EPA Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AD62969 - General Preparation										
Duplicate (AD62969-DUP1)	Source: 16D0202-01			Prepared: 04/06/16 Analyzed: 04/08/16						
Total Suspended Solids	468	1.0	mg/L		490			4.70	30	
Duplicate (AD62969-DUP2)	Source: 16D0361-01			Prepared: 04/06/16 Analyzed: 04/08/16						
Total Suspended Solids	382	1.0	mg/L		335			13.2	30	
Batch AD63089 - General Prep										
Blank (AD63089-BLK1)	Prepared & Analyzed: 04/11/16									
Phosphorus, total	ND	0.040	mg/L							
LCS (AD63089-BS1)	Prepared & Analyzed: 04/11/16									
Phosphorus, total	0.200	0.040	mg/L	0.200		99.8	85-115			
Duplicate (AD63089-DUP1)	Source: 16D0370-02			Prepared & Analyzed: 04/11/16						
Phosphorus, total	3.21	0.80	mg/L		3.24			0.831	20	
Matrix Spike (AD63089-MS1)	Source: 16D0370-02			Prepared & Analyzed: 04/11/16						
Phosphorus, total	4.96	0.80	mg/L	2.00	3.24	86.1	70-130			
Matrix Spike (AD63089-MS2)	Source: 16D0372-02			Prepared & Analyzed: 04/11/16						
Phosphorus, total	5.04	0.80	mg/L	2.00	3.51	76.7	70-130			
Matrix Spike Dup (AD63089-MSD1)	Source: 16D0370-02			Prepared & Analyzed: 04/11/16						
Phosphorus, total	4.99	0.80	mg/L	2.00	3.24	87.4	70-130	0.543	20	
Batch AD63127 - General Preparation										
Blank (AD63127-BLK1)	Prepared: 04/06/16 Analyzed: 04/11/16									
Biochemical Oxygen Demand	ND	5.0	mg/L							



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Conventional Chemistry Parameters by APHA/EPA Methods - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC Limits	RPD	RPD Limit	Flag
Batch AD63127 - General Preparation									
Blank (AD63127-BLK2)				Prepared: 04/06/16 Analyzed: 04/11/16					
Biochemical Oxygen Demand	ND	5.0	mg/L						
LCS (AD63127-BS1)				Prepared: 04/05/16 Analyzed: 04/10/16					
Biochemical Oxygen Demand	220	5.0	mg/L	200	110	84-115			
Duplicate (AD63127-DUP1)				Source: 16D0344-02 Prepared: 04/06/16 Analyzed: 04/11/16					
Biochemical Oxygen Demand	ND	5.0	mg/L		ND			30	



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Anions by EPA Method 300.0 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
Batch AD62920 - General Preparation										
Blank (AD62920-BLK1)				Prepared & Analyzed: 04/05/16						
Nitrite as N	ND	0.20	mg/L							
Nitrate as N	ND	0.20	mg/L							
Nitrite as NO2	ND	1.0	mg/L							
Nitrate as NO3	ND	1.0	mg/L							
Chloride	ND	0.50	mg/L							
LCS (AD62920-BS1)				Prepared & Analyzed: 04/05/16						
Nitrite as NO2	18.0	1.0	mg/L	18.2		98.8	90-110			
Nitrate as NO3	26	1.0	mg/L	24.7		104	90-110			
Nitrate as N	5.80	0.20	mg/L	5.56		104	90-110			
Nitrite as N	5.47	0.20	mg/L	5.56		98.5	90-110			
Chloride	11.5	0.50	mg/L	11.1		103	90-110			
Duplicate (AD62920-DUP1)				Source: 16D0166-01		Prepared & Analyzed: 04/05/16				
Chloride	3.78	0.50	mg/L		3.78			0.0264	20	
Nitrate as NO3	ND	1.0	mg/L		ND				20	
Nitrite as N	ND	0.20	mg/L		ND				20	
Nitrate as N	ND	0.20	mg/L		ND				20	
Nitrite as NO2	ND	1.0	mg/L		ND				20	
Matrix Spike (AD62920-MS1)				Source: 16D0166-01		Prepared & Analyzed: 04/05/16				
Chloride	14.9	2.5	mg/L	11.1	3.78	100	80-120			
Nitrite as N	5.30	1.0	mg/L	5.56	ND	95.4	80-120			
Nitrate as N	5.44	1.0	mg/L	5.56	ND	98.0	80-120			
Nitrite as NO2	17.4	5.0	mg/L	18.2	ND	95.7	80-120			
Nitrate as NO3	24	5.0	mg/L	24.7	ND	97.7	80-120			
Matrix Spike (AD62920-MS2)				Source: 16D0167-01		Prepared & Analyzed: 04/06/16				
Nitrate as N	7.01	0.20	mg/L	5.56	0.237	122	80-120			QM-07
Nitrite as NO2	20.1	1.0	mg/L	18.2	ND	110	80-120			
Nitrate as NO3	31	1.0	mg/L	24.7	1.0	122	80-120			QM-07
Nitrite as N	6.12	0.20	mg/L	5.56	ND	110	80-120			
Chloride	14.7	0.50	mg/L	11.1	3.13	105	80-120			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Anions by EPA Method 300.0 - Quality Control

Analyte(s)	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Flag
------------	--------	--------------------	-------	----------------	------------------	------	----------------	-----	--------------	------

Batch AD62920 - General Preparation

Matrix Spike Dup (AD62920-MSD1)

Source: 16D0166-01

Prepared & Analyzed: 04/05/16

Chloride	14.9	2.5	mg/L	11.1	3.78	100	80-120	0.112	20	
Nitrate as N	5.47	1.0	mg/L	5.56	ND	98.4	80-120	0.437	20	
Nitrite as NO2	17.5	5.0	mg/L	18.2	ND	96.1	80-120	0.418	20	
Nitrate as NO3	24	5.0	mg/L	24.7	ND	98.2	80-120	0.437	20	
Nitrite as N	5.32	1.0	mg/L	5.56	ND	95.8	80-120	0.418	20	



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

West Bay Sanitary District
500 Laurel Street
Menlo Park, CA 94025

Project Manager: John Simonetti
Project: WQ Monitoring
Project Number: Oak Avenue

Reported:
04/19/16 10:42

Notes and Definitions

- QM-01 The spike recovery for this QC sample is outside of established control limits possibly due to a sample matrix interference.
- QM-05 The spike recovery was outside acceptance limits for the MS and/or MSD due to matrix interference. The LCS and/or LCSD were within acceptance limits showing that the laboratory is in control and the data is acceptable.
- QM-07 The spike recovery was outside acceptance limits for the MS and/or MSD. The batch was accepted based on acceptable LCS recovery.
- QM-4X The spike recovery was outside of QC acceptance limits for the MS and/or MSD due to analyte concentration at 4 times or greater the spike concentration. The QC batch was accepted based on LCS and/or LCSD recoveries within the acceptance limits.
- T-1 This sample was received outside recommended holding time.
- T-14 Residual chlorine, dissolved oxygen, and pH must be analyzed in the field to meet the EPA specified 15 minute hold time.
- T-16 Sample prepared outside of recommended hold time. Client notified.
- T-2 Sample analyzed outside of recommended holding time per client.
- ND Analyte NOT DETECTED at or above the reporting limit
- dry Sample results reported on a dry weight basis
- REC Recovery
- RPD Relative Percent Difference

Company/Name: West Bay Sanitary District		Project ID: WQ Monitoring		Signature below authorizes work under terms stated on reverse side.												Sample Notes		TAT											
Mailing Address: 500 Laurel Street Menlo Park, CA 94025		Project No: Oak Avenue		Analysis Request												4.9		24 hr <input type="radio"/> 48 hr <input type="radio"/> Lab Approval Required <input type="radio"/> 1 wk <input type="radio"/> 2 wk (standard) <input type="radio"/>											
Project Contact & Email (Hardcopy or PDF to): John Simonetti jsimonetti@westbaysanitary.org		PO#																											
Phone/Fax:		Site Contact:																											
Field Sampler - Print Name & Signature		Sample Collection		Container		Preservative		Matrix																					
Sample Identification		Date Time		40ml Vial	Poly	Glass	Sleeve	Other	HCL	HNO3	H2SO4	Other	None	Water	Soil	Other	DU BOD	DU pH	200.7 - B, Ca, Mg, Na	Ammonia, Phosphorus	Total Nitrogen	Chloride, EC, NO2, NO3	Alkalinity, Bicarbonate	SAR Adjusted Calculation	Silica	TDS	TSS		
Oak Avenue		7-1-16 0900			4					1	1		2	x			x	x	x	x	x	x	x	x	x	x	x		